



SEMTRONIK

MICRONAIRE TESTER 009
Digital Micronaire Fibre Fineness System



MICRONAIRE TESTER MODEL SMNT 009 Digital Micronaire Fibre Fineness System

Micronaire is a critical cotton fiber quality that is harder to visualize than other properties like staple length, fiber strength or colour. It is more influenced by weather than just about any other fiber characteristic.

Micronaire describes both the maturity and fineness of cotton fiber. The end quality of cotton yarn or fabric is partially determined by micronaire.

When cotton's micronaire is too low, nep and dyeability problems result during textile processing.

When micronaire measures higher and closer to the discount range, it cannot be as efficiently spun into high quality, fine-count yarns

Cotton's quality is determined by the structure and dimensions of its fibers. Staple, length and micronaire are the dimensions of fiber.

Micronaire has no units of measurement, but is generally understood to be the weight in micrograms per inch of fiber length. Intrinsic fiber strength is thought to be influenced mainly by the structure of the primary and secondary cell walls of the fibers.

The commercial HVI measure of fiber bundle strength is called tenacity, which is the intrinsic strength per unit linear density and thus is directly influenced by one of the fiber dimensions, that is, micronaire.

To be specific, HVI strength is calculated from the breaking strength of fiber bundles in grams force to break divided by the micronaire value expressed in tex units.

Essentially, tex value is the metric equivalent of micronaire, or micrograms per millimeter of fiber length.

Micronaire is reported to be influenced to a greater extent by environmental conditions than by genetics. While this may be generally true, it by no means predicates that variety selection cannot have an important impact on the micronaire of the crop.

Micronaire is a very sensitive measurement, and a small change can make a big difference.

For example, the difference between a micronaire of 4 and a micronaire of 5 is only 1 microgram per inch of fiber. There are a million micrograms in a gram and approx. 454 grams in a pound.

Such a minute difference can amount to a major difference when it comes to the farmer's bottom line.

The system is equipped with a sensor to measure the micronaire, interfaced with necessary electronics to display the value locally. There is also a facility to communicate the value to the PC using RS-232.

System is calibrated with CIRCOT calibration cotton.

The system also consists of a weigh scale with ISI approval and accuracy of up to 3 decimal places.

The system is calibrated according to ASTM standards.

TECHNICAL DATA

Application Range

Measurement of cotton micronaire
Fibre material (100 % Cotton)
Bale, roller and saw ginned.

Measuring principle: Air flow

Measurements: Micronaire value.
Range from approx. 2.4 to 6.5

Sample size: 9.97 to 10.03 grams
Auto eject after measuring process.

Duration of measurement:
Approx. 25 seconds for two (2)
repetitions per sample.

Operation:

- Minimum operation training required.
- Pneumatic operation.
- Function keys on the front panel of the instrument.
- Data presentation by LCD display on the front panel of instrument.
- Self calibration check.

Connection possibilities:

Connection to a PC system for data collection and analysis.

Dimensions:

W 410 x H 305 x D 600 mm
Weight: Approx. 23 kg (50 lbs)

Power:


Voltage: 230 VAC
Frequency: 50 Hertz
Consumption: 250 Watts
UPS 0.75/1KVA included.

Compressed Air:

Air pressure 4 to 6 bar.
(60 to 90 PSI)

Weigh Scale:

Range: 0 to 100 gms.
Resolution: 0.01 gms.
ISI Certified.



For further details, contact



Website: www.semitronik.com

E-mail: semiahd@satyam.net.in

Since our policy is of continuous development and improvement, we reserve the right to supply product which may differ from those illustrated & described in this publication.

HEAD OFF.
17 CD,
Archana Ind. Estate,
Rakhlal Road,
Ahmedabad-380 023.
Tel.: 079 22741011,
22742480, 22774977.
Fax: 079 22741793.

MUMBAI OFF.
B12, Kasturchand
Mill Estate,
M. C. Jawie Marg,
Dadar (W),
Mumbai-400 028.
Tel.: 022 24221485.
Fax: 022 24322755.

SURAT OFF.
D. K. Vasanwala
407, Trade Centre,
Ring Road,
Surat-395 002.
Tel.: 0261 2354847.
Fax: 0261 2324746.
M: 9374722631

ERODE OFF.
Randhir Singh
146, Meenakshi
Sunderam Street,
Thrunagar Colony,
Erode.
M: 9344035935

AMRITSAR OFF.
House No. 6 /A,
Mata Kaulan Marg,
Kashmir Avenue,
Amritsar-144 001.
M: 9356124229

KOLKATA OFF.
52, Kamdhenu,
5TH Floor,
4A, Ray Street,
Bhawaniipur,
Kolkata-700 020.

LUDHIANA OFF.
M: 09868905214